

# VERIFICATION OF CALIBRATION REPORT

OCT 04 2010

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130135</u> ✓	
Name <u>Trp. J. Santora</u>		ID# <u>2276</u> ✓	Date <u>9/22/10</u> ✓
A	Agency <u>Alaska State Troopers</u>		Phone # <u>656-1233</u>
Instrument Location <u>Galena Police Department</u>			
Alco S/N <u>78855</u> ✓ <u>130135</u> ✓ <u>10/6/10</u>		Target Value <u>.081</u> ✓	High Pressure <u>1075</u> (tank pressure)
B	Alco Test Values 1 <sup>st</sup> Alco <u>.081</u> ✓		2 <sup>nd</sup> Alco <u>.080</u> ✓
Signature <u>[Signature]</u>			
(OVER)			


I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 16<sup>th</sup> day of DEC, 2010.

[Signature] (Notary Seal Stamp)  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



CONTACTED TRPR SANTORA 10/5/10  
AND VERIFIED ALCO BOTTLE SERIAL  
NUMBER MATCHED 78855 AS  
INDICATED BY SCOL: DATABASE ON  
10/6/10

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N \_\_\_\_\_ 130135 \_\_\_\_\_

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130135

SEPTEMBER 22, 2010

OPERATOR'S NAME:

SANTORA/JEANNINE/L

OPERATOR'S NUMBER: 2276

SUBJECT'S LAST NAME:

VERIFICATION

SUBJECT'S FIRST NAME/MI :

OF CAL

O.L. #: 0000000

DEPT/AGENCY: GALI

CASE/REPORT:

TEST TYPE: V

ALCO TARGET VALUE: .001

ALCO S/N: 130135-78855 10/6/10

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.90 in  
ALCO TARGET .000 14:58  
BLANK TEST .000 14:59  
INTERNAL STANDARD VERIFIED 14:59  
ALCO TO 29.90 in .001 15:00  
BLANK TEST .000 15:01  
SUBJECT SAMPLE .000 15:02  
BLANK TEST .000 15:03  
ALCO TO 29.90 in .000 15:03  
BLANK TEST .000 15:04

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130135

SEPTEMBER 22, 2010

TIME 15:06

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 29.90 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

"#22'()++-./0123456789:;<=>@AECDEFG  
HIJKL MNOPQRSTUVWXYZ[-\]^\_`abcde fghi jklmno  
pqrstuvwxyz{|}~`"